

Certificate of Analysis

CERTIFIED REFERENCE MATERIAL

RD0261061.5: KIT of 21 components - part 2 of 3;
7 components; 10ug/ml each of Prometryn [CAS:7287-19-6] ; Terbutryn [CAS:886-50-0] ; Terbutylazine-desethyl [CAS:30125-63-4] ; Prometon [CAS:1610-18-0] ; Desmetryn [CAS:1014-69-3] ; Atrazine - desethyl - desisopropyl [CAS:3397-62-4] ; Ametryn [CAS:834-12-8] in Acetone

Lot N: 756552
Barcode: 92761506

Ref N: RD0264781.5

Certification Date: 14.06.2021

Component	Certified Value* and uncertainty [µg/ml]	CAS	Chemical Formula
Prometryn	9.984 ± 0.111	7287-19-6	C ₁₀ H ₁₉ N ₅ S
Terbutryn	9.983 ± 0.219	886-50-0	C ₁₀ H ₁₉ N ₅ S
Terbutylazine-desethyl	10.054 ± 0.178	30125-63-4	C ₇ H ₁₂ ClN ₅
Prometon	9.988 ± 0.190	1610-18-0	C ₁₀ H ₁₉ N ₅ O
Desmetryn	10.015 ± 0.143	1014-69-3	C ₈ H ₁₅ N ₅ S
Atrazine - desethyl - desisopropyl	10.034 ± 0.180	3397-62-4	C ₃ H ₄ ClN ₅
Ametryn	10.023 ± 0.126	834-12-8	C ₉ H ₁₇ N ₅ S

* WQP 5.15.1/2 The certified value was obtained gravimetrically and confirmed experimentally by GC/MS or HPLC

Density 0.7865 g/cm³ at 20°C

Starting Material	Purity, Batch
Prometryn	99.9% (41401842)
Terbutryn	99.9% (41398647)
Terbutylazine-desethyl	97.4% (41356746)
Prometon	98.3% (41359716)
Desmetryn	99.59% (41377222)
Atrazine - desethyl - desisopropyl	94.5% (41365045)
Ametryn	97.9% (41395189)

Storage Conditions: Store in a refrigerator at temperatures between 2°C to 8°C

Expiry Date: 14.07.2022

Concept of Certification and traceability statement:

This certified reference material is produced by gravimetric measurement and dissolving the individual substances in Acetone .
The reported expanded uncertainty of measurement is stated as the standard uncertainty of measurement multiplied by the coverage factor $k = 2$, which for a normal distribution corresponds to a coverage probability of approximately 95%. The standard uncertainty of measurement has been determined in accordance with EA 4/02 and incorporates the uncertainties of the raw-material purity, the mass and the volume.
The metrological traceability is defined as the "property of a measurement result whereby the result can be related to a reference through a documented unbroken chain of calibrations, each contributing to the measurement uncertainty".
The metrological traceability is ensured through gravimetric measurement and dissolving of certified reference material/s (traceable to SI) from laboratories/producers, accredited according to ISO 17034.
The measurement results are traceable to SI. All analytical balances used for the preparation of the solution are calibrated yearly under an in-house procedure with class E1 and class E2 analytical weights, traceable to SI (DKD), and are checked daily.
Class A laboratory glassware is used.
The results from temperature measurement are traceable to SI. The thermometers used for solution's calibration are calibrated from an ISO 17025 accredited laboratory. The ambient conditions are controlled with a hygrometer calibrated from an ISO 17025 accredited laboratory.
Both, purity of the starting materials and solvent, were checked using appropriate analytical instrument.

Intended use: For Laboratory Use Only

This CRM is intended for:



C.P.A. chem Ltd is accredited to ISO 17034 and ISO/IEC 17025

CPAchem Ltd
www.cpacchem.com
e-mail: info@cpacchem.com; tel.: +359 42 60 77 16
for France, Belgium and Switzerland:
e-mail: acsd2@wanadoo.fr; tel.: 01 30 57 57 32 / fax: 01 30 57 57 33

Calibration of TLC, GC/FID, GC/TCD, GC/ECD, GC/MS, GC/MS/MS, LC/UV, LC/MS and LC/MS/MS
 Validation of analytical methods
 Preparation of "working reference samples"
 Detection limit and linearity studies
 This statement is not intended to restrict the use for other purposes.

Instructions for the correct use of this certified reference material:

This CRM can be used directly or can be diluted in an appropriate solvent. Only a clean class A glassware should be used. Do not pipet from container. Obtained concentration (in mg/l) after dilution is a result from the multiplication of certified value of CRM concentration and the CRM's volume used for dilution and divided into the flask's volume used for dilution. For quantitative analysis, we recommend analyzing this mixture separately, without mixing it with other solutions, to ensure accurate results for every compound.

Stability and storage:

This CRM is with a guaranteed stability until $\pm 5\%$ of the certified concentration for a period of 12 months. Stability is guaranteed of an unopened original packaging stored, as written in the section: Storage Conditions. Even if the product is stable at normal laboratory conditions, in order to increase its stability, we highly recommend it to be stored in a refrigerator. The product should be used shortly after opening to avoid concentration changes due to evaporation. Warranty does not apply to a product stored after opening.

Hazardous situation:

The normal laboratory safety precautions should be observed when working with this RM. Further details for the handling of this RM are available in a safety data sheet.

Level of homogeneity

This solution was mixed according to an in-house procedure (MQP 5.13.1) and is guaranteed to be homogeneous. To ensure sufficient homogeneity of the sample prior to use thoroughly mix by inversion or sonicate.

Names of certifying officers:

Laboratory:  Margarita Dimitrova

Manager:  Krassimira Taralova

This document QF 5.17.1/1 version 1 is designed and the certified value(s) and uncertainty(ies) are determined in accordance with ISO Guide 31, ISO Guide 35, and Eurachem / CITAC Guides

This certificate relates solely to the lot number given above.

All processes (including generating of this certificate) are completely controlled by the specialized Computer-Aided-Manufacturing (CAM) software.

This Certified Reference Material was produced under a quality management system that is:

- Registered to ISO 9001 Quality Management System (Lloyd's Register Quality Assurance Ltd Cert No 0039638)
- Accredited according to ISO/IEC 17025 – Testing (ANAB Cert No AT-1836)
- Accredited according to ISO 17034 - Reference Material Producer (ANAB Cert No AR-1835)

Additional Information

Gravimetric Data

Component	Purity %	Source Lot No	Weighed quantity, g	Final quantity, $\text{kg} \cdot 10^{-3}$	Bulk/ Standard Solution lot No	Concentration mg/kg	Chemist ID
Prometryn	99.9	41401842	0.03544	3.5781	91739100	9894.8	AS
		91739100	0.1512	11.8697	92761926	126.044	ER
		92761926	1.5842	15.7293	92761506	12.6947	ER
Terbutryn	99.9	41398647	0.00939	2.8325	91739469	3311.8	AS
		91739469	0.4517	11.8697	92761926	126.030	ER
		92761926	1.5842	15.7293	92761506	12.6933	ER
Terbutylazine-desethyl	97.4	41356746	0.01202	2.9071	91740137	4027.2	AS
		91740137	0.3741	11.8697	92761926	126.926	ER
		92761926	1.5842	15.7293	92761506	12.7835	ER
Prometon	98.3	41359716	0.01137	2.5696	91719829	4349.6	AS
		91719829	0.3441	11.8697	92761926	126.096	ER
		92761926	1.5842	15.7293	92761506	12.6999	ER
Desmetryn	99.59	41377222	0.03128	3.5682	91739940	8730.4	AS
		91739940	0.1719	11.8697	92761926	126.436	ER

		92761926	1.5842	15.7293	92761506	12.7341	ER
Atrazine - desethyl - desisopropyl	94.5	41365045	0.0125	2.6939	91731036	4384.9	AS
		91731036	0.3429	11.8697	92761926	126.673	ER
		92761926	1.5842	15.7293	92761506	12.7580	ER
Ametryn	97.9	41395189	0.02732	3.1094	91739094	8601.8	AS
		91739094	0.1746	11.8697	92761926	126.531	ER
		92761926	1.5842	15.7293	92761506	12.7437	ER

