

CERTIFIED REFERENCE MATERIAL

15 components; 1000mg/l each of Aniline [CAS:62-53-3] ; Diphenylamine [CAS:122-39-4] ; o-Toluidine [CAS:95-53-4] ; o-Anisidine [CAS:90-04-0] ; 3-Anisidine [CAS:536-90-3] ; 4-Anisidine [CAS:104-94-9] ; p-Toluidine [CAS:106-49-0] ; Nitrobenzene [CAS:98-95-3] ; 1,2-Dinitrobenzene [CAS:528-29-0] ; 1,3-Dinitrobenzene [CAS:99-65-0] ; 1-Chloro-2-nitrobenzene [CAS:88-73-3] ; 1-Chloro-3-nitrobenzene [CAS:121-73-3] ; 1-Chloro-4-nitrobenzene [CAS:100-00-5] ; 2,5-Dichloronitrobenzene [CAS:89-61-2] ; 3,4-Dichloronitrobenzene [CAS:99-54-7] in Methanol

Lot N: 766706
Barcode: 92789944

Ref N: RD0593031.3

Certification Date: 22.08.2021

Component	Certified Value* and uncertainty [µg/ml]	CAS	Chemical Formula
Aniline	1020.1 ± 9.9	62-53-3	C ₆ H ₇ N
Diphenylamine	1003.0 ± 20.4	122-39-4	C ₁₂ H ₁₁ N
o-Toluidine	1008.0 ± 9.8	95-53-4	C ₇ H ₉ N
o-Anisidine	1016.0 ± 9.8	90-04-0	C ₇ H ₉ NO
3-Anisidine	1004.6 ± 9.7	536-90-3	C ₇ H ₉ NO
4-Anisidine	1023.0 ± 20.4	104-94-9	C ₇ H ₉ NO
p-Toluidine	1014.0 ± 20.4	106-49-0	C ₇ H ₉ N
Nitrobenzene	1004.1 ± 9.7	98-95-3	C ₆ H ₅ NO ₂
1,2-Dinitrobenzene	1007.0 ± 20.4	528-29-0	C ₆ H ₄ N ₂ O ₄
1,3-Dinitrobenzene	1013.9 ± 20.9	99-65-0	C ₆ H ₄ N ₂ O ₄
1-Chloro-2-nitrobenzene	1021.0 ± 20.4	88-73-3	C ₆ H ₄ ClNO ₂
1-Chloro-3-nitrobenzene	1003.0 ± 20.4	121-73-3	C ₆ H ₄ ClNO ₂
1-Chloro-4-nitrobenzene	1010.0 ± 20.4	100-00-5	C ₆ H ₄ ClNO ₂
2,5-Dichloronitrobenzene	1013.0 ± 20.4	89-61-2	C ₆ H ₃ Cl ₂ NO ₂
3,4-Dichloronitrobenzene	1007.9 ± 20.3	99-54-7	C ₆ H ₃ Cl ₂ NO ₂

* WQP 5.15.1/2 The certified value was obtained gravimetrically and confirmed experimentally by GC/MS or HPLC

Density 0.8087 g/cm³ at 20°C

Starting Material	Purity, Batch
Aniline	99.9% (41390498)
Diphenylamine	99.9% (41427941)
o-Toluidine	99.7% (41415238)
o-Anisidine	99.5% (41385883)
3-Anisidine	98.0% (41430866)
4-Anisidine	99.9% (41392195)
p-Toluidine	99.9% (41417669)
Nitrobenzene	99.9% (41419274)
1,2-Dinitrobenzene	99.9% (41398258)
1,3-Dinitrobenzene	99.5% (41171882)
1-Chloro-2-nitrobenzene	99.9% (41391235)
1-Chloro-3-nitrobenzene	99.9% (41387726)
1-Chloro-4-nitrobenzene	99.8% (41425114)
2,5-Dichloronitrobenzene	99.9% (41430804)
3,4-Dichloronitrobenzene	99.1% (41393697)

Storage Conditions: Store in a freezer at -18°C and protected from light



Expiry Date: 22.03.2022

Concept of Certification and traceability statement:

This certified reference material is produced by gravimetric measurement and dissolving the individual substances in Methanol. The reported expanded uncertainty of measurement is stated as the standard uncertainty of measurement multiplied by the coverage factor $k = 2$, which for a normal distribution corresponds to a coverage probability of approximately 95%. The standard uncertainty of measurement has been determined in accordance with EA 4/02 and incorporates the uncertainties of the raw-material purity, the mass and the volume. The metrological traceability is defined as the "property of a measurement result whereby the result can be related to a reference through a documented unbroken chain of calibrations, each contributing to the measurement uncertainty". The metrological traceability is ensured through gravimetric measurement and dissolving of certified reference material/s (traceable to SI) from laboratories/producers, accredited according to ISO 17034. The measurement results are traceable to SI. All analytical balances used for the preparation of the solution are calibrated yearly under an in-house procedure with class E1 and class E2 analytical weights, traceable to SI (DKD), and are checked daily. Class A laboratory glassware is used. The results from temperature measurement are traceable to SI. The thermometers used for solution's calibration are calibrated from an ISO 17025 accredited laboratory. The ambient conditions are controlled with a hygrometer calibrated from an ISO 17025 accredited laboratory. Both, purity of the starting materials and solvent, were checked using appropriate analytical instrument.

Intended use: For Laboratory Use Only

This CRM is intended for:
Calibration of TLC, GC/FID, GC/TCD, GC/ECD, GC/MS, GC/MS/MS, LC/UV, LC/MS and LC/MS/MS
Validation of analytical methods
Preparation of "working reference samples"
Detection limit and linearity studies
This statement is not intended to restrict the use for other purposes.

Instructions for the correct use of this certified reference material:

This CRM can be used directly or can be diluted in an appropriate solvent. Only a clean class A glassware should be used. Do not pipet from container. Obtained concentration (in mg/l) after dilution is a result from the multiplication of certified value of CRM concentration and the CRM's volume used for dilution and divided into the flask's volume used for dilution. For quantitative analysis, we recommend analyzing this mixture separately, without mixing it with other solutions, to ensure accurate results for every compound.

Stability and storage:

This CRM is with a guaranteed stability until $\pm 5\%$ of the certified concentration for a period of 6 months. Stability is guaranteed of an unopened original packaging stored, as written in the section: Storage Conditions. Even if the product is stable at normal laboratory conditions, in order to increase its stability, we highly recommend it to be stored in a refrigerator. The product should be used shortly after opening to avoid concentration changes due to evaporation. Warranty does not apply to a product stored after opening.

Hazardous situation:

The normal laboratory safety precautions should be observed when working with this RM. Further details for the handling of this RM are available in a safety data sheet.

Level of homogeneity

This solution was mixed according to an in-house procedure (MQP 5.13.1) and is guaranteed to be homogeneous. To ensure sufficient homogeneity of the sample prior to use thoroughly mix by inversion or sonicate.

Names of certifying officers:

Laboratory:  Margarita Dimitrova

Manager:  Krassimira Taralova

This document QF 5.17.1/1 version 1 is designed and the certified value(s) and uncertainty(ies) are determined in accordance with ISO Guide 31, ISO Guide 35, and Eurachem / CITAC Guides

This certificate relates solely to the lot number given above.

All processes (including generating of this certificate) are completely controlled by the specialized Computer-Aided-Manufacturing (CAM) software.

This Certified Reference Material was produced under a quality management system that is:

- Registered to ISO 9001 Quality Management System (Lloyd's Register Quality Assurance Ltd Cert No 0039638)
- Accredited according to ISO/IEC 17025 – Testing (ANAB Cert No AT-1836)
- Accredited according to ISO 17034 – Reference Material Producer (ANAB Cert No AR-1835)

Additional Information

Gravimetric Data

Component	Purity %	Source Lot No	Weighed quantity, g	Final quantity, $\text{kg} \cdot 10^{-3}$	Bulk/ Standard Solution lot No	Concentration mg/kg	Chemist ID
Aniline	99.9	41390498	0.102	8.1024	92789883	12576.3	NN

		92789883	0.8111	8.0866	92789944	1261.42	NN
Diphenylamine	99.9	41427941	0.01004	8.0866	92789944	1240.32	NN
o-Toluidine	99.7	41415238	0.101	8.1024	92789883	12428.0	NN
		92789883	0.8111	8.0866	92789944	1246.56	NN
o-Anisidine	99.5	41385883	0.102	8.1024	92789883	12525.9	NN
		92789883	0.8111	8.0866	92789944	1256.37	NN
3-Anisidine	98.0	41430866	0.1024	8.1024	92789883	12385.5	NN
		92789883	0.8111	8.0866	92789944	1242.28	NN
4-Anisidine	99.9	41392195	0.01024	8.0866	92789944	1265.03	NN
p-Toluidine	99.9	41417669	0.01015	8.0866	92789944	1253.91	NN
Nitrobenzene	99.9	41419274	0.1004	8.1024	92789883	12379.0	NN
		92789883	0.8111	8.0866	92789944	1241.63	NN
1,2-Dinitrobenzene	99.9	41398258	0.01008	8.0866	92789944	1245.26	NN
1,3-Dinitrobenzene	99.5	41171882	0.01019	8.0866	92789944	1253.81	NN
1-Chloro-2-nitrobenzene	99.9	41391235	0.01022	8.0866	92789944	1262.56	NN
1-Chloro-3-nitrobenzene	99.9	41387726	0.01004	8.0866	92789944	1240.32	NN
1-Chloro-4-nitrobenzene	99.8	41425114	0.01012	8.0866	92789944	1248.96	NN
2,5-Dichloronitrobenzene	99.9	41430804	0.01014	8.0866	92789944	1252.68	NN
3,4-Dichloronitrobenzene	99.1	41393697	0.01017	8.0866	92789944	1246.32	NN

